



**Batumi Oil Terminal LLC
Measuring Equipment Calibration Laboratory
Scope of Accreditation**



ССТ ИСО/ИЭК 17025:2010
GAC-TL-0087
25.11.2014 – 25.11.2018

**Batumi Oil Terminal LLC
Measuring Equipment Calibration Laboratory
Scope of Accreditation**



Batumi Oil Terminal LLC
Measuring Equipment Calibration Laboratory
Scope of Accreditation



ССТ ИСО/ИЭК 17025:2010
 GAC-TL-0087
 25.11.2014 – 25.11.2018

№№	Calibration object	Scope of measurement	Calibration and measuring abilities	Method identification and title
1.	Reference Manometer	Pressure	(0-40) kgf/sm ² , Accuracy class 0,4	BOT-CL-3-0005 Calibration of reference manometers with 0.4 accuracy class
2.	Digital Pressure Calibrator	Pressure	(0-20) kgf/sm ² , Accuracy class 0,4	BOT-CL-3-0005 Calibration of reference manometers with 0.4 accuracy class
3.	Technical Manometer, Mano-vacuum meter (pressure)	Pressure	(0-40) kgf/sm ² , Accuracy class 1,0 - 2,5	BOT-CL-3-0003 Calibration of technical manometers / mano-vacuum meters (pressure)
4.	Digital Thermometer with RTD sensor	Temperature	(0...+100) °C, Accuracy ±1.5 °C	BOT-CL-3-0021 Method of comparison of digital electronic thermometers
5.	Bimetallic Thermometer / Manometer	Temperature	(0...+100) °C, Accuracy class 1,0	BOT-CL-3-0022 Calibration of bimet/ manometric thermometers